

Certificate of Analysis

Research Use Only - Not for Use in Diagnostic Procedures

Product	Material Code	Batch Number
ACCURUN® SARS-CoV-2 Antigen Reference Material Kit	2015-0231	10542972

Product Description:

Kit Components	Material Code	Batch Number
Positive Reference Material	2015-0229	10542970
Negative Reference Material	2015-0230	10542971

Date of Manufacture: 24 FEB 2021

Expiration Date: 24 AUG 2022

Storage Temperature: -10 to -25 °C

QC Testing:

Kit Component	Test	Qualitative Result
SARS-CoV-2 Antigen Positive	Roche cobas Elecsys® SARS-CoV-2 Antigen	Positive
	Ortho VITROS® SARS-CoV-2 Antigen	Positive
SARS-CoV-2 Antigen Negative	Roche cobas Elecsys® SARS-CoV-2 Antigen	Negative
	Ortho VITROS® SARS-CoV-2 Antigen	Negative

Safety Precautions: Universal transport media manufactured with plasma negative for HIV1 and 2, HCV, HBsAg and non-reactive for anti-bodies to SARS-CoV-2. Use the Centers for Disease Control recommended universal precautions.

Approval:

Ch. McNeill

21 SEP 2021

Prepared By

Date

Amy L. Lane

21 SEP 2021

QA Verified By

Date