

Certificate of Analysis


Product Description: ACCURUN Swab Flu A/B Reference Material
 Material Number: 2020-0178
 Batch Number: 10684944
 Storage Temperature: Ambient
 Date of Manufacture: 15 NOV 2023
 Date of Expiration: 15 NOV 2025
 QC Testing:

Testing was performed using Cepheid Xpert® Xpress CoV-2/Flu/RSV plus assay run on the GeneXpert system and ID NOW™ Influenza A & B 2 assay performed on the ID NOW Instrument. The Cepheid instrument detects 2 targets for Flu A: Flu A1 and Flu A2 and both Ct values are given in the table below.

Component	Test Method	Result
ACCURUN Swab Flu A/B Reference Material	Cepheid	Flu A Positive Flu A1 = 29.18 Ct Flu A2 = 31.76 Ct Flu B Positive 31.74 Ct
	Abbott ID NOW	Flu A Positive Flu B Positive

Safety Precautions: Preparation contains heat inactivated Flu A/California/04/2009 H1N1 and Flu B/Florida/78/2015 Victoria viruses. Use the Center for Disease Control recommended universal precautions.

Approval:


 Prepared By _____ Date 12 DEC 2023

 QA Verified By _____ Date 13 DEC 2023