

# Certificate of Analysis

Product Description: ACCURUN® HCV LOD Reference Material Kit  
 Material Number: 2020-0211  
 Batch Number: 10714561  
 Storage Temperature: -70°C  
 Date of Manufacture: 02JUL2024  
 Date of Expiration: 02JUL2027

Testing was performed on the Roche cobas 5800/6800/8800 system using the Roche cobas HCV assay.

## QC Testing

| Product                                       | Test Method              | Result                       |
|---|--------------------------|------------------------------|
| ACCURUN® HCV<br>LOD Reference<br>Material Kit | Roche cobas HCV<br>assay | 20/20 replicates<br>positive |

## Safety Precautions:

Preparation contains live Hepatitis C virus derived from donor material.  
 Use the Center for Disease Control recommended universal precautions.

## Approval:



28AUG2024

Prepared By

Date



QA Verified By

26 Sept 2024  
Date