

Certificate of Analysis

Product Description: ACCURUN® HIV-1 LOD Reference Material Kit
 Material Number: 2020-0212
 Batch Number: 10714562
 Storage Temperature: -70°C
 Date of Manufacture: 02JUL2024
 Date of Expiration: 02JUL2027

Testing was performed on the Roche cobas 5800/6800/8800 system using the Roche cobas HIV-1 assay.

QC Testing

Product	Test Method	Result
ACCURUN® HIV-1 LOD Reference Material Kit	Roche cobas HIV-1 assay	20/20 replicates positive

Safety Precautions:

Preparation contains a non-replicable HIV-1 8E5 strain. Use the Center for Disease Control recommended universal precautions.

Approval:



28AUG2024

Prepared By

Date



QA Verified By

Date

24 Sept 2024