

Certificate of Analysis

Product Description: ACCURUN® CMV LOD Reference Material Kit
Material Number: 2020-0227
Batch Number: 10755238
Storage Temperature: -70°C
Date of Manufacture: 11 JUN 2025
Date of Expiration: 19 MAR 2027

Testing was performed on the Roche cobas 5800/6800/8800 system using the Roche cobas CMV assay.

QC Testing:

Product	Test Method	Results
ACCURUN CMV LOD Reference Material Kit	Roche cobas CMV assay	≥ 95% of replicates detected

Safety Precautions:

Preparation contains whole organism CMV in a plasma-based diluent containing 0.09% sodium azide. Use the Center for Disease Control recommended universal precautions.

Approval:

	06 OCT 2025
Prepared By	Date
	06 OCT 2025
QA Verified By	Date