

Certificate of Analysis

Product Description: ACCURUN® CMV 1E3 IU/mL Precision Kit
Material Number: 2020-0228
Batch Number: 10780473
Storage Temperature: -70°C
Date of Manufacture: 19 SEP 2025
Date of Expiration: 26 AUG 2027

Testing was performed on the Roche cobas 5800/6800/8800 system using the Roche cobas CMV assay. The measured concentration is the average of the 10 replicates tested.



QC Testing:

Product	Test Method	Result
ACCURUN CMV 1E3 IU/mL Precision Kit	Roche cobas CMV assay	Measured Concentration: 961 IU/mL

Safety Precautions:

Preparation contains whole organism CMV in a plasma-based diluent containing 0.09% sodium azide. Use the Center for Disease Control recommended universal precautions.

Approval:

	03 OCT 2025
Prepared By	Date
	03 Oct 2025
QA Verified By	Date