

Certificate of Analysis

Product Description: ACCURUN® CMV 5E5 IU/mL Precision Kit
 Material Number: 2020-0229
 Batch Number: 10780474
 Storage Temperature: -70°C
 Date of Manufacture: 17 SEP 2025
 Date of Expiration: 18 AUG 2027

Testing was performed on the Roche cobas 5800/6800/8800 system using the Roche cobas CMV assay. The measured concentration is the average of the 10 replicates tested.

QC Testing:

Product	Test Method	Result
ACCURUN CMV 5E5 IU/mL Precision Kit	Roche cobas CMV assay	Measured Concentration: 4.58E+05 IU/mL

Safety Precautions:

Preparation contains whole organism CMV in a plasma-based diluent containing 0.09% sodium azide. Use the Center for Disease Control recommended universal precautions.

Approval:


 Prepared By

03 OCT 2025
 Date


 QA Verified By

03 Oct 2025
 Date