

## **Certificate of Analysis**

For Research Use Only Not for Human or Animal Consumption

## KPL SureBlue<sup>™</sup> TMB Microwell Peroxidase Substrate (1-Component)

PART CODE: LOT NUMBER: SIZE	5120-0078 (52-00-04) <b>10583101</b> 5000 mL
DESCRIPTION:	This liquid substrate system develops a dark blue soluble product when reacted with horseradish peroxidase labeled conjugates in microwell plates or tubes. It is not recommended for membrane or immunohistochemical staining assays.
FORM/STORAGE:	Liquid. Store refrigerated at 2 - 8°C.
CONTENT:	The one component system is supplied as 5000 mL bottles, from bulk Lot No <b>. 10583101</b> . The KPL SureBlue™ TMB Microwell Peroxidase Substrate (1-Component) contains 3,3',5,5'- tetramethyl-benzidine in an acidic buffer. The concentration of the Hydrogen Peroxide is 0.01%.
STABILITY:	Product is stable for three years from date of manufacture when stored properly.
DATE OF MANUFACTURE:	29 OCT 2021
<b>Approval:</b>	<ul> <li>KPL SureBlue<sup>™</sup> TMB Microwell Peroxidase Substrate (1-Component) is extremely sensitive to certain handling and storage conditions. The following suggestions are provided to ensure success when redispensing this product:</li> <li>Avoid prolonged exposure to light air or extreme temperatures; however, exposure to indirect laboratory lighting for 24 hours is not harmful.</li> <li>Never pipette directly from bottle. Pour out needed amount into a plastic reservoir. Do not return excess KPL SureBlue<sup>™</sup> TMB to the primary container.</li> <li>Avoid exposure to silica-based materials including borosilicate glass, metals, bacterial contamination or other oxidizing agents. Do not refilter as this may destabilize the dye complex and result in the development of background color.</li> <li>Redispense into amber Nalgene HDPE or LDPE bottles ONLY. Do not redispense into glass bottles as this may compromise substrate performance.</li> <li>HDPE and LDPE bottles should be utilized as they are received directly from the manufacturer; washing of bottles is not recommended.</li> <li>Redispense substrate using Cole-Parmer C-Flex tubing and do not reuse tubing. Avoid redispensing with all other tubings including silicone, latex, and Tygon.</li> <li>Gravity fill or the use of a peristaltic pump is advised.</li> <li>Avoid substrate contact with any metallic surfaces.</li> <li>For further details concerning the redispensing of this product, please contact the Technical Service Dept. at 800-638-3167.</li> </ul>
Approvai:	

New alciereii 12 NOV 202	1
QA Verified By Date	