

CERTIFICATE OF ANALYSIS



BacTrace[®] Shigella Genus Mix Positive Control Antigen

CATALOG NO. 50-90-01
LOT NO. 090845
SIZE 1.0 mL

DESCRIPTION Heat-killed *Shigella boydii*, *flexneri*, *dysenteriae* mix in dextran solution.

**LOT SPECIFIC
CELL COUNT** **3 x 10⁹ cfu/ml

FORM/STORAGE Lyophilized. Store refrigerated at 2 – 8°C.

**STABILIZERS AND
PRESERVATIVES** Dextran added to stabilize during lyophilization. No preservatives. Non-sterile.

STABILITY Product is stable for one year from date of receipt when stored properly.

**DATE OF
MANUFACTURE** 08/27/2009