

# Certificate of Analysis

## KPL H. pylori Positive Control Antigen

**PART Code** 5370-0012 (50-93-94)  
**Lot No.** 10518844  
**SIZE** 1.0 mL

**DESCRIPTION** Heat-killed *Helicobacter pylori* cells (ATCC strain 43504), 1 mL of approximately 1% wet packed cells w/v in dextran solution.

**LOT SPECIFIC CELL COUNT** Cell count =  $4 \times 10^8$  CFU/mL

**FORM/STORAGE** Lyophilized. Store refrigerated at 2 – 8°C or ambient room temperature.

**STABILIZERS AND PRESERVATIVES** Dextran added to stabilize during lyophilization. No preservatives. Non-sterile.

**STABILITY** Product is stable for one year from date of receipt when stored properly.

**DATE OF MANUFACTURE** 27 MAY 2021

Approval:

A handwritten signature in black ink, appearing to read "Kimberly Smith".

Quality Assurance

09 JUL 2021

Date