



CERTIFICATE OF ANALYSIS

H. pylori Positive Control Antigen

CATALOG NO. 50-93-94
LOT NO. 150464
SIZE 1.0 mL

DESCRIPTION Heat-killed *Helicobacter pylori* cells (ATCC strain 43504), 1 mL of approximately 1% wet packed cells w/v in dextran solution.

LOT SPECIFIC CELL COUNT Cell count = 1.65×10^8 cells/mL

FORM/STORAGE Lyophilized. Store refrigerated at 2 – 8°C or ambient room temperature.

STABILIZERS AND PRESERVATIVES Dextran added to stabilize during lyophilization. No preservatives. Non-sterile.

STABILITY Product is stable for one year from date of receipt when stored properly.

DATE OF MANUFACTURE 08/04/2015

Quality Assurance Verified and Approved:

A handwritten signature in black ink, appearing to read 'Christina Quinonez', is written over a white background.

Date: 08/14/2015