

www.seracare.com

Certificate of Analysis

KPL 10X Tris-Glycine-SDS Buffer

PART Code 5460-0016 (51-11-02)

Lot No. 10275934

SIZE 5 L

DESCRIPTION Gel electrophoresis buffer. To prepare 1X running buffer for Tris-Glycine-SDS gel

electrophoresis, mix 100 mL KPL 10X Tris-Glycine-SDS buffer and 900 mL reagent quality

water. Do not adjust pH.

FORM/STORAGE Liquid. Store at room temperature.

CONTENTS Supplied as a 5L container(s).

STABILITY Product is stable for one year from date of receipt when stored at room temperature.

DATE OF

MANUFACTURE 24 JUL 2017

Approval:

31 JUL 2017

Quality Assurance

Date

910 Clopper Road, Gaithersburg, MD 20878 • 508.244.6400 • 800.676.1881 Toll Free • 508.634.3334 Fax