

Manufacturing Quality  
ISO 9001:2015  
ISO 13485:2016

## Certificate of Analysis

**Onco-Ref™ Genomic DNA Reference Standard EGFR c.2390G>C/c.2369C>T  
p.C797S/T790M COSM5945664/6240**

### Product Information

*serivane Batch 291023  
Drucker 26MARCH2019*

<b>Catalog Number:</b>	ASE-6176G-1
<b>Lot Number:</b>	E6176G291023-1
<b>Expiration Date:</b>	2020-09-11
<b>Quantity/Volume/Size:</b>	1 µg / 20 µl / 1 vial
<b>Gene Name:</b>	EGFR
<b>Nucleotide Change:</b>	c.2390G>C/c.2369C>T
<b>Amino Acid Change:</b>	p.C797S/T790M
<b>Mutation Id:</b>	COSM5945664/6240
<b>Cell:</b>	RKO
<b>Storage Temperature:</b>	-20°C
<b>Shipping Temperature:</b>	Ambient
<b>Instructions for Use:</b>	Refer to product Datasheet (Available <a href="http://www.accuref.com">www.accuref.com</a> )
<b>Intended Use:</b>	Research Use Only. Not Intended for human or animal diagnostic or therapeutic use.

### Quality Control

Test	Method	Acceptance Criteria	Result
Genotyping	Sanger Sequencing	Homozygous	PASS
Quality	Gel Electrophoresis	No Smear	PASS
Quantification	Qubit 4.0 Fluorometer	50 -60 ng /µl	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

<i>Deepika Tewari</i>	<b>2018-10-01</b>
<b>Quality Assurance</b>	<b>Date</b>

COA-ASE-6176G-1-V2