

Manufacturing Quality

ISO 9001:2015

ISO 13485:2016

Certificate of Analysis

Onco-Ref™ Genomic DNA Reference Standard PDGFRA c.2524_2535del12
 p.I843_D846delIMHD COSM737


Product Information

Catalog Number:	ASE-6256G-1
Lot Number:	G256200122
Expiration Date:	2022-01-21
Quantity/Volume/Size:	1 µg / 20 µl / 1 vial
Gene Name:	PDGFRA
Nucleotide Change:	c.2524_2535del12
Amino Acid Change:	p.I843_D846delIMHD
Mutation Id:	COSM737
Cell:	HCT116
Storage Temperature:	-20°C
Shipping Temperature:	Ambient
Instructions for Use:	Refer to product Datasheet (Available www.accuref.com)
Intended Use:	For Research Use Only. Not intended for human or animal diagnostics or therapeutics.

Quality Control

Test	Method	Acceptance Criteria	Result
Genotyping	Sanger Sequencing	Homozygous	PASS
Quality	Gel Electrophoresis	No Smear	PASS
Quantification	Qubit 4.0 Fluorometer	50 -60 ng /µl	PASS
DNA Purity	Nanodrop	A260/280 Ratio Between 1.8 to 2	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

	2020-01-29
Quality Assurance	Date

COA-ASE-6256G-1-V2