



Manufacturing Quality

ISO 9001:2015

## **Certificate of Analysis**

ISO 13485:2016

## Onco-Ref<sup>™</sup> Genomic DNA Reference Standard TP53 c.533\_534insN p.H178fs\*3 COSM44361

## **Product Information**

| Catalog Number:       | ASE-6519G-1  |  |  |
|-----------------------|--|--|--|
| Lot Number:           | G519191009   |  |  |
| Expiration Date:      | 2021-10-08   |  |  |
| Quantity/Volume/Size: | 1 μg / 20 μl / 1 vial  |  |  |
| Gene Name:            | TP53   |  |  |
| Nucleotide Change:    | c.533_534insN  |  |  |
| Amino Acid Change:    | p.H178fs*3   |  |  |
| Mutation Id:          | COSM44361  |  |  |
| Cell:                 | HCT116   |  |  |
| Storage Temperature:  | -20°C  |  |  |
| Shipping Temperature: | Ambient  |  |  |
| Instructions for Use: | Refer to product Datasheet (Available www.accuref.com)                             |  |  |
| Intended Use:         | Research Use Only. Not Intended for human or animal diagnostic or therapeutic use. |  |  |

## **Quality Control**

| Test           | Method                | Acceptance Criteria | Result |
|----------------|-----------------------|---------------------|--------|
| Quality        | Gel Electrophoresis   | No Smear            | PASS   |
| Quantification | Qubit 4.0 Fluorometer | 50 -60 ng /µl       | PASS   |
| Genotype-LS    | Sanger Sequencing     | Locus Specific PCR  | PASS   |

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

| King Muye         | 2019-10-17 |
|-------------------|------------|
| Quality Assurance | Date       |
|                   |            |

COA-ASE-6519G-1-V2