

Manufacturing Quality

ISO 9001:2015

ISO 13485:2016

Certificate of Analysis

Onco-Ref™ FFPE Cell Slide Reference Standard BRAF 1799_1800TG>AA
 V600E COSM475


Product Information

Catalog Number:	ASO-3046
Lot Number:	O3046171214
Expiration Date:	2020-12-13
Quantity/Volume/Size:	5 µm thick section with 7.0mm diameter cell core / 1 pack (20 slides)
Gene Name:	BRAF
Nucleotide Change:	1799_1800TG>AA
Amino Acid Change:	V600E
Mutation Id:	COSM475
Cell:	HCT116
Storage Temperature:	-20°C
Shipping Temperature:	Ambient
Instructions for Use:	Refer to product Datasheet (Available www.accuref.com)
Intended Use:	Research Use Only. Not Intended for human or animal diagnostic or therapeutic use.

Quality Control

Test	Method	Acceptance Criteria	Result
Genotyping	Sanger Sequencing	Homozygous	PASS
Quality (H&E)	H&E Staining	Stain Quality	PASS
Cell Density	Visual	>60%	PASS

AccuRef Diagnostics certifies that this standard meets the specifications stated in this certificate and warrants this product to meet the stated acceptance criteria through the expiration date when stored unopened as recommended.

	2019-05-30
Quality Assurance	Date

COA-ASO-3046-V1