

# Certificate of Analysis

**Product Description:** Toxoplasma IgG Positive Plasma  
**Material No(s):** 0320-0003, 0320-0004, 0320-0005, 0320-0006  
**Lot No:** BM149290  
**Donor ID:** NA  
**Gender:** NA  
**Age:** NA  
**Country of Origin:** NA  
**Collection Date:** 05-Aug-02  
**Anticoagulant:** NA  
**Storage Temperature:** -20 degrees C  
**Shipping Conditions:** Dry ice  
**Intended Use:** For research use only. Not for use in diagnostic procedures.  
**Safety Precautions:** Use universal precautions when handling this material.  
**Test Method(s):** See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
20-Aug-07	Toxoplasma IgG	Abbott AxSym	9.2	S/CO	27.6 IU/mL
22-Aug-07	Toxoplasma IgG	Diamedix EIA Quantitative	5	Index	250 IU/mL

This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HIV Ag, Syphilis, and found to be negative.

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
16-Aug-07	Toxoplasma IgM	Trinity Biotech Captia EIA	0.484	S/CO	NA
22-Aug-07	Toxoplasma IgM	Diamedix EIA	0.263	S/CO	NA
30-Aug-07	Toxoplasma IgM	Abbott AxSym	0.325	S/CO	0.195 Index

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**

*Lynne Labonte*

29-Sep-15

Prepared By

*[Signature]*

Date

02-Oct-15

QA Verified By

Date

