

Certificate of Analysis

Product Description: Zika IgG Positive Plasma
 Material No(s): 0325-0072, 0325-0073, 0325-0074
 Lot No: 10219009
 Donor ID: BD303746
 Gender: Male
 Age: 43
 Country of Origin: US
 Collection Date: 14-Jul-16
 Anticoagulant: K2 EDTA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
23-Aug-16	Zika IgG	Euroimmun EIA	31.162	RU/ML	>22 RU/ML is Positive
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA, WNV RNA, Syphilis and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
23-Aug-16	Zika IgM	Euroimmun EIA	0.022	S/CO	NA
01-Sep-16	Dengue IgG	Calbiotech EIA	2.653	S/CO	NA
01-Sep-16	Dengue IgM	Calbiotech EIA	0.266	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Rabente

15-Sep-16

Prepared By

Date

Luva

15-Sep-16

QA Verified By

Date