

Certificate of Analysis

Product Description: Zika IgG Positive Plasma
 Material No(s): 0325-0072, 0325-0073, 0325-0074
 Lot No: 10219021
 Donor ID: BD303758
 Gender: Male
 Age: 55
 Country of Origin: US
 Collection Date: 13-Jul-16
 Anticoagulant: K2 EDTA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

| Lot Analysis | | | | | |
|---|----------|--------------------------|--------|-------|-----------------------|
| Test Date | Analyte | Test Method/Manufacturer | Result | Units | Notes |
| 23-Aug-16 | Zika IgG | Euroimmun EIA | 48.178 | RU/ML | >22 RU/ML is Positive |
| This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA, WNV RNA, Syphilis and found to be negative. | | | | | |

| Additional Data | | | | | |
|-----------------|------------|--------------------------|--------|-------|-------|
| Test Date | Analyte | Test Method/Manufacturer | Result | Units | Notes |
| 23-Aug-16 | Zika IgM | Euroimmun EIA | 0.026 | S/CO | NA |
| 01-Sep-16 | Dengue IgG | Calbiotech EIA | 2.824 | S/CO | NA |
| 01-Sep-16 | Dengue IgM | Calbiotech EIA | 0.403 | S/CO | NA |

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Rabente

15-Sep-16

Prepared By

Date

Buwa

15-Sep-16

QA Verified By

Date