

AccuSpan[™] HCV RNA Linearity Panel

2410-0166 / Batch #10682172

OVERVIEW


The AccuSpan[™] HCV RNA Linearity Panel (2410-0166 / Batch #10682172) is an eight-member panel made from serial dilutions of high titer HCV positive plasma with established reactivity for HCV (Hepatitis C) RNA. This panel consists of seven members representing serial log dilutions of HCV positive plasma in HCV RNA negative diluent and one negative member prepared from the diluent. The diluent was prepared from normal human plasma that was filtered through a 0.2 micron filter. Sodium azide (0.09%) was added as a preservative.

Results are reported for each panel member on specific test methods. Linearity is shown graphically by plotting observed results against expected results. The WHO International Standard was tested in the same run as the AccuSpan HCV RNA Linearity Panel members. Both expected and observed results for the standards are reported; expected results for the WHO standards are based upon application of a dilution factor to the WHO assigned value.

For Research Use Only. Not for use in diagnostic procedures. Data are offered for informational purposes. LGC SeraCare Life Sciences does not claim that others can duplicate test results exactly.

CAUTION: These materials have not been treated and should be considered biohazardous. Follow Universal Precautions.

AccuSpan[™] HCV RNA Linearity Panel



HCV RNA results were obtained using the Roche COBAS[®] AmpliPrep/COBAS[®] TaqMan[®] HCV Test, v2.0 test method. Results are the mean of three replicates. A line of best fit is shown.

AccuSpan™ HCV RNA Linearity Panel

2410-0166 / Batch #10682172

HCV RNA^{1,2}

Panel Member	Roche COBAS® AmpliPrep/COBAS® Taqman® HCV Test, v2.0	Abbott m2000 Realtime HCV
	(log IU/mL)	(log IU/mL)
01	6.85³	6.90³
02	5.86	5.84
03	4.92	4.86
04	4.00	3.81
05	3.23	2.88
06	2.10	1.99
07	<1.18, TND, <1.18*	<1.08, <1.08, 1.15*
08	TND	TND
Test Date	15-Sep-2023	06-Sep-2023
Test Site	RL	RL
Test Kit Range	15 to 100,000,000 IU/mL 1.18 to 8.00 log IU/mL	12 to 100,000,000 IU/mL 1.08 to 8.00 log IU/mL
Kit Part Code	NA	NA
Kit Lot No.	K04281	382108
Kit Exp. Date	30-Apr-2024	30-Apr-2024

¹Results are reported as log international units per mL (log IU/mL); positive/reactive results are noted in bold red.

²Results are reported as the mean result of triplicate testing.

³Panel member #1 was tested at a 1:10 dilution and results were corrected for dilution factor.

*Interpretation variability based on triplicate testing, individual replicates reported.

TND = Target Not Detected; RL = Reference Lab; NA = Not available

AccuSpan™ HCV RNA Linearity Panel

2410-0166 / Batch #10682172

6th WHO International HCV RNA Standard (18/184)

Panel Member	Expected Values (log IU/mL)	Observed Values on Roche COBAS® AmpliPrep/COBAS® TaqMan® HCV Test, v2.0	% Difference ²
		(log IU/mL) ¹	
01	4.70	4.52	-3.9
02	4.00	3.86	-3.4
03	3.70	3.53	-4.5
04	3.00	2.89	-3.7
Test Date		15-Sep-2023	
Test Site		RL	
Test Kit Range		15 to 100,000,000 IU/mL 1.18 to 8.00 log IU/mL	
Kit Part Code		NA	
Kit Lot No.		K04281	
Kit Exp. Date		30-Apr-2024	

¹WHO panel was tested in the same test run as the AccuSpan™ HCV RNA Linearity Panel members. Samples were run in singlet. Positive/reactive results are noted in bold red.

²Percentage difference is how much the observed concentration differs from the expected concentration. Values calculated for reference only. Laboratories may use the data to apply a correction factor to the test results.
RL = Reference Lab; NA = Not available

The package insert for this panel can be found at www.seracare.com.

A printed copy of the package insert or data sheet may be requested by email at CDx-Info@LGCGroup.com or by phone at 508.244.6400.