



# ACCURUN® 106 SERIES 1000

HIV-1 Antigen Positive  
Quality Control

## About this package insert

Thank you for your interest in this ACCURUN product.

This package insert consists of two pages.

The first page contains the product name and an explanation of the symbols used on the labeling.

The second page contains the complete package insert text.

If the package insert you view or print does not contain two pages, or if you experience any problems, please email us at [info@seracare.com](mailto:info@seracare.com).

By phone: US customers call 800.676.1881; International customers call collect 508.634.3359.

A printed package insert will be sent to you upon request.



MEDIMARK® Europe  
11, rue Émile Zola BP 2332  
38033 Grenoble Cedex 2 – France  
+ 33 (0) 4 76 86 43 22  
[info@medimark-europe.com](mailto:info@medimark-europe.com)



Manufacturer: SeraCare Life Sciences, Inc. | 25 Birch Street, Milford, MA 01757 USA  
Phone: 001.508.244.6400 | [info@seracare.com](mailto:info@seracare.com)

March 2013 12022US-04

## Explanation of symbols used in SeraCare product labeling



Harmful/Irritant

This product contains  
0.1% ProClin® 300.

R43 May cause sensitization by  
skin contact.

S24 Avoid contact with skin.

S35 This material and its  
container must be disposed  
of in a safe way.

S37 Wear suitable gloves.



Upper limit  
of temperature



Biological risks



Negative control



Positive control



Temperature  
limitation



Use By



Catalogue  
number



Batch code



"Caution, consult  
accompanying documents"



Authorized Representative in  
the European Community



*In Vitro* Diagnostic  
Medical Device



**THESE REAGENTS MUST NOT BE SUBSTITUTED FOR THE MANDATORY POSITIVE AND NEGATIVE CONTROL REAGENTS PROVIDED WITH MANUFACTURED TEST KITS.**

#### NAME AND INTENDED USE

ACCURUN controls are intended to estimate laboratory testing precision and can be used to detect errors in laboratory testing procedures. ACCURUN<sup>®</sup> 106 HIV-1 Antigen Positive Quality Control Series 1000 is formulated for use with *in vitro* diagnostic test methods that detect HIV-1 antigen. *For In Vitro Diagnostic Use.*

#### SUMMARY

Frequent testing of independent quality control samples provides the analyst with a means of monitoring the performance of laboratory assays. Routine use of controls enables laboratories to monitor day-to-day test variation, lot-to-lot performance of test kits, and operator variation, and can assist in identifying increases in random or systematic error. A well designed quality control program can provide added confidence in the reliability of results obtained for unknown specimens. The use of independent controls may provide valuable information concerning laboratory proficiency and kit lot variation that may affect assay sensitivity<sup>1</sup>.

#### PRINCIPLES OF THE PROCEDURE

ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 is designed for use with *in vitro* assay procedures for the purpose of monitoring test performance. ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 is manufactured from HIV-1 viral lysate derived from human cell culture and is diluted in human serum or plasma nonreactive for HBsAg and negative for antibodies to HIV 1 and 2, HCV and HTLV. ACCURUN controls do not have assigned values. Specific levels of reactivity will vary among different manufacturer's assays, different procedures, different lot numbers and different laboratories.

#### REAGENTS

Cat. No. A106-1003            3 vials, 5.0 mL per vial  
Cat. No. A106-1008            12 vials, 3.5 mL per vial

This control contains stabilizers (EDTA, buffering agents) and 0.1% ProClin<sup>®</sup> (5-chloro-2-methyl-4-isothiazolin-3-one & 2-methyl-4-isothiazolin-3-one) as preservative. The HIV-1 viral lysate has been inactivated by treatment with Triton X-100, beta propiolactone, and ultraviolet irradiation.

#### WARNINGS AND PRECAUTIONS

##### *For In Vitro Diagnostic Use*

**CAUTION:** Handle ACCURUN controls and all human blood products as though capable of transmitting infectious agents. ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 is manufactured from HIV-1 viral lysate derived from human cell culture, and is diluted in human serum or plasma nonreactive for HBsAg and negative for antibodies to HIV 1 and 2, HCV and HTLV with current FDA licensed tests. Infectious materials have been inactivated with Triton X-100, beta propiolactone, and ultraviolet irradiation.

##### Safety Precautions

Use the Centers for Disease Control (CDC) recommended universal precautions for handling ACCURUN and human blood<sup>2</sup>. Do not pipette by mouth; do not smoke, eat or drink in areas where specimens are being handled. Clean any spillage by immediately wiping up with 0.5% sodium hypochlorite solution. Dispose of all specimens, controls, and materials used in testing as though they contain infectious agents.

##### Handling Precautions

Do not use ACCURUN controls beyond the expiration date. Avoid contamination of the controls when opening and closing the vials.

##### STORAGE INSTRUCTIONS

Store ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 at 2-8°C. Once opened, ACCURUN controls should be discarded after 60 days. After opening, record the date opened and the expiration date on the vial. Multiple freeze-thaw cycles are not recommended, and may have variable adverse effects upon test results. To prevent leakage, store vials upright.

##### INDICATIONS OF REAGENT INSTABILITY OR DETERIORATION

Alterations in physical appearance may indicate instability or deterioration of ACCURUN controls. Solutions that are visibly turbid should be discarded.

##### PROCEDURE

##### Materials Provided

ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 is manufactured from HIV-1 viral lysate derived from human cell culture and human serum or plasma nonreactive for HBsAg and negative for antibodies to HIV 1 and 2, HCV and HTLV.

##### Materials Required but not Provided

Refer to instructions supplied by manufacturers of the test kits to be used.

#### Instructions for Use

Allow the controls to reach room temperature prior to use, then return controls to refrigerated storage immediately after use. Mix the contents of the vials by gentle inversion. ACCURUN controls should be included in a test run using exactly the same procedure provided by the manufacturer for unknown specimens. ACCURUN controls must NOT be substituted for the positive and negative control reagents provided with manufactured test kits.

#### Quality Control

Since ACCURUN controls do not have assigned values, it is recommended that each laboratory validate the use of each lot of ACCURUN with each specific assay system prior to its routine use in the laboratory.

#### INTERPRETATION OF RESULTS

Levels of reactivity of ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 may vary with different manufacturers' tests and different test kit lots. Since the control does not have an assigned value, the laboratory must establish a range for each lot of ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000. When results for ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 are outside the established acceptance range of values, it may be an indication of unsatisfactory test performance. Possible sources of error include: deterioration of test kit reagents, operator error, faulty performance of equipment, or contamination of reagents.

#### LIMITATIONS OF THE PROCEDURE

**ACCURUN CONTROLS MUST NOT BE SUBSTITUTED FOR THE POSITIVE AND NEGATIVE CONTROL REAGENTS PROVIDED WITH MANUFACTURED TEST KITS.**

TEST PROCEDURES and INTERPRETATION OF RESULTS provided by manufacturers of test kits must be followed closely. Deviations from procedures recommended by test kit manufacturers may produce unreliable results. ACCURUN controls are not calibrators and should not be used for assay calibration. Performance characteristics for ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 have been established only for HIV-1 antigen. Adverse shipping and storage conditions or use of outdated controls may produce erroneous results.

#### EXPECTED RESULTS

**ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 DOES NOT HAVE AN ASSIGNED VALUE.** This control is formulated to produce positive reactivity with the test kits listed in Table 1. Specific levels of reactivity will vary among different manufacturers' assays, different procedures, different lot numbers and different laboratories. Procedures for implementing a quality assurance program and monitoring test performance on a routine basis must be established by each individual laboratory. Each laboratory should establish its own range of acceptable values. For example, the acceptable range might include all values within 2 standard deviations of the mean of 20 data points obtained in 20 runs over a period of 30 days<sup>3</sup>.

#### SPECIFIC PERFORMANCE CHARACTERISTICS

ACCURUN controls have been designed for use with *in vitro* assay procedures for purposes of monitoring assay performance. ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 is formulated to be reactive for HIV-1 antigen and is diluted in human serum or plasma that is tested and found nonreactive for HBsAg and negative for antibodies to HIV 1 and 2, HCV, and HTLV. ACCURUN controls do not have assigned values. Specific levels of reactivity will vary among different manufacturers' assays, different procedures, different lot numbers, and different laboratories. Procedures for implementing a quality assurance program and monitoring test performance on a routine basis must be established by each individual laboratory.

#### REFERENCES

- Green IV GA, Carey RN, Westgard JO, Carten T, Shablesky LA, Achord D, Page E, and Le AV. *Quality control for qualitative assays: quantitative QC procedure designed to assure analytical quality required for an ELISA for hepatitis B surface antigen.* Clin. Chem. 43:9 1618-1621, 1997.
- CDC *recommendations for prevention of HIV transmission in health care settings.* MMWR 36 (supp. 2), 1987.
- Statistical Quality Control for Quantitative Measurements: Principles and Definitions; Approved Guideline—Second Edition.* NCCLS document C24-A2, 1999.

**Table 1.** ACCURUN 106 HIV-1 Antigen Positive Quality Control Series 1000 is formulated to produce positive reactivity with the following test kit:

Manufacturer	Assay
Abbott Diagnostics, Sligo Ireland	Architect HIV Ag/Ab Combo

**For assistance, contact SeraCare Technical Support at 001.508.244.6400.**