

# Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*

Product Description: Seraseq® ctDNA Reference Material v2 WT

Material Number: 0710-0208      Batch Number: 10282172

Material Description: ctDNA-like human genomic DNA from the reference cell line, GM24385.

Concentration (Qubit dsDNA BR Assay): 30.7 ng/mL

Volume: 5.0 mL

Date of Manufacture: 10 Aug 2017      Expiration Date: 10 Aug 2020

Storage: 2-8°C

# Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*

Product Description: Seraseq® ctDNA Reference Material v2 WT

Gene ID	COSMIC Identifier	AF% <sup>1</sup>
AKT1	COSM33765	0.03
APC	COSM13127	0.05
APC	COSM18561	0.00
ATM	COSM21924	0.00
BRAF	COSM476	0.00
CTNNB1	COSM5664	0.02
EGFR	COSM6224	0.02
EGFR	COSM12378	0.02
EGFR	COSM6225	0.00
EGFR	COSM6240	0.02
ERBB2	COSM20959	0.00
FGFR3	COSM715	0.00
FLT3	COSM783	0.00
FOXL2	COSM33661	0.00
GNA11	COSM52969	0.00
GNAQ	COSM28758	0.00
GNAS	COSM27887	0.06
IDH1	COSM28747	0.08
JAK2	COSM12600	0.00
KIT	COSM1314	0.01
KRAS	COSM521	0.02
MPL	COSM18918	0.00
NCOA4-RET	N/A	0.00
NPM1	COSM17559	0.00
NRAS	COSM584	0.09
PDGFRA	COSM736	0.03
PDGFRA	COSM28053	0.00
PIK3CA	COSM763	0.03
PIK3CA	COSM12464 <sup>3</sup>	0.00
PIK3CA	COSM775	0.10
PTEN	COSM4986	0.00
PTEN	COSM5809	0.00
RET	COSM965	0.01
SMAD4	COSM14105	0.01
TP53	COSM10648	0.05
TP53	COSM10660	0.03
TP53	COSM10662	0.07
TP53	COSM6530	0.00
TP53	COSM18610	Not Tested <sup>2</sup>
TPR-ALK	N/A	0.01

Digital PCR testing using  
BioRad QX200™ Droplet  
Digital™ PCR System:

<sup>1</sup> Variant allele frequencies > 0.00% for this wildtype negative control are within the expected range for stochastic positive dPCR reactions.

<sup>2</sup> NGS testing alone is used for QC of variant COSM18610.

<sup>3</sup> As of June 2019 this mutation is no longer listed in the COSMIC database.

# Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*

Product Description: Seraseq® ctDNA Reference Material v2 WT

NGS was performed as an orthogonal verification step. Results confirm no variants were detected above 0.1%.

**NGS Parameters:**

DNA input = 50 ng  
# of samples / flow cell = 3  
# of total reads / sample = 5.6 M  
Average read depth = 4506X  
On-target reads = ~97%  
Q30 score = ~92%  
Analysis = Archer Analysis Suite v5.1.1.1

Next Generation Sequencing testing using Archer® Reveal ctDNA™ 28 Kit run on an Illumina® MiSeq™ using v2 (2x150 bp) PE chemistry reagents<sup>1</sup>:

Default analysis settings except for the following:

- ERROR\_CORRECTION is turned to ON (default=OFF; cfDNA pipeline)
- NORMALIZATION\_LEVEL is set to 10000000 (default is 3500000; cfDNA pipeline)
- VARIANT\_DOWNSTREAM\_ROI\_SIZE is set to 150 (default=400; cfDNA pipeline)
- BARCODE\_HAMMING\_DISTANCE=1 (default=2)

<sup>1</sup>Please see the poster from NIST for more information about assay sensitivity:

<https://digital.seracare.com/multilab-assessment-reference-materials-ctdna-poster2018>

Approval:



Prepared By

13 Jun 2019

Date