

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® ctDNA Complete Reference Material AF 2.5%

Material Number: 0710-0670 Batch Number: 10437716

Material Description: A ctDNA-like mixture of human genomic DNA from the reference cell line, GM24385, and synthetic DNA constructs

Concentration (Qubit dsDNA BR Assay): Nominal value: 25 ng/mL; Average measured value after extraction using Qiagen QIAamp Circulating Nucleic Acid Kit: 30.4 ng/mL

Fill Volume: 5.0 mL

Date of Manufacture: 05 AUG 2019 Expiration Date: 05 AUG 2022

Storage: 2-8°C

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® ctDNA Complete Reference Material AF 2.5%

Digital PCR testing using
BioRad QX200™ Droplet
Digital™ PCR System:

Gene ID	COSMIC Identifier	Amino Acid Change	Average AF%
AKT1	COSM33765	p.E17K	2.63
BRAF	COSM476	p.V600E	2.59
EGFR	COSM6224	p.L858R	2.61
EGFR	COSM6240	p.T790M	2.49
ERBB2	COSM20959	p.A775_G776insYVMA	2.27
KIT	COSM1314	p.D816V	2.68
KRAS	COSM521	p.G12D	2.79
NCOA4/RET	NA	Translocation	2.51
NRAS	COSM584	p.Q61R	2.85
PIK3CA	COSM775	p.H1047R	2.54
PIK3CA	COSM12464 ¹	p.N1068fs*4	2.54
EML4-ALK	NA	Translocation	2.47
ALK	COSM144250	p.G1202R	2.37
ALK	COSM28055	p.F1174L	2.37
BRCA1	COSM1383519	p.K654fs*47	2.29
BRCA2	COSM1738242	p.R2645fs*3	2.39
EGFR	COSM12370	p.L747_P753>S	3.04
EGFR	COSM6256	p.S752_I759delSPKANKEI	2.51
EGFR	COSM6223	p.E746_A750delELREA	2.97
KRAS	COSM516	p.G12C	2.69
CD74/ROS1	NA	Translocation	2.61
KRAS	COSM554	p.Q61H	2.35

Gene ID	Average CNV in ctDNA ²	Average Additional Copies (per cell) in ctDNA
ERBB2	4.64	2.64
MET	3.72	1.72
MYC	3.93	1.93

NA = not applicable

¹As of June 2019, this mutation is no longer listed in the COSMIC database.

²Compare to a normal CNV of 2.00.

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® ctDNA Complete Reference Material AF 2.5%

Next Generation Sequencing testing using Archer® Reveal ctDNA™ 28 Kit run on an Illumina® MiSeq™ using v2 (2x150 bp) PE chemistry reagents^{1,2}:

Gene ID	COSMIC Identifier	Amino Acid Change	AF%
AKT1	COSM33765	p.E17K	1.98
BRAF	COSM476	p.V600E	3.05
EGFR	COSM6224	p.L858R	2.26
EGFR	COSM6240	p.T790M	3.02
ERBB2	COSM20959	p.A775_G776insYVMA	1.80
KIT	COSM1314	p.D816V	2.67
KRAS	COSM521	p.G12D	2.98
NCOA4/RET	NA	Translocation	NA
NRAS	COSM584	p.Q61R	2.57
PIK3CA	COSM775	p.H1047R	2.64
PIK3CA	COSM12464 ³	p.N1068fs*4	2.18
EML4-ALK	NA	Translocation	NA
ALK	COSM144250	p.G1202R	1.98
ALK	COSM28055	p.F1174L	2.27
BRCA1	COSM1383519	p.K654fs*47	NA
BRCA2	COSM1738242	p.R2645fs*3	NA
EGFR	COSM12370	p.L747_P753>S	3.25
EGFR	COSM6256	p.S752_I759delSPKANKEI	3.08
EGFR	COSM6223	p.E746_A750delELREA	3.13
KRAS	COSM516	p.G12C	1.95
CD74/ROS1	NA	Translocation	NA
KRAS	COSM554	p.Q61H	2.48

Gene ID	CNV in ctDNA ⁴	Additional Copies (per cell) in ctDNA
ERBB2	4.16	2.16
MET	4.54	2.54
MYC	NA	NA

NA = not applicable; AF% and CNV marked NA were not targeted by the panel.

¹NGS was performed as an orthogonal verification step. Parameters used:

- DNA input = 50 ng
- # of samples / flow cell = 3
- # of total reads / sample = 1.8M
- Average read depth = 5111X
- On-target reads = 94.6%
- Q30 score = 92.7%

Analysis = Archer Analysis Suite v5.1.7 (default settings except for: Error correction was on, MAPQ threshold for variant call was 10, minimum allele fraction for variant call of

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® ctDNA Complete Reference Material AF 2.5%

0.00025, minimum base quality for variant call of 30)

²Please see the poster from NIST for more information about assay sensitivity:

<https://digital.seracare.com/multilab-assessment-reference-materials-ctdna-poster2018>

³As of June 2019, this mutation is no longer listed in the COSMIC database.

⁴Compare to a normal CNV of 2.00.

Note: The MET gene is amplified using two synthetic constructs with a small region of overlap between the constructs (see package insert for genomic coordinates). Assays which target this region of overlap may report higher amplification levels.

Approval:



Prepared By

07 AUG 2019

Date