



# Quality Control Certificate

**Product Description:** Optitrol Yellow

**Catalog Number:** SR11021

**Lot:** 10809736 (Box)

10809735 (Vial)

**Expiry Date:** 31 MAR 2028

**Preservative:** Sodium Azide; ProClin® 950

**Application:** For *in vitro* diagnostic use

**Storage:** 2 to 8 °C

**Human Viral Marker:** The controls are prepared from processed human plasma or serum reactive for antibodies to HIV 1, antibodies to HTLV I, antibodies to HCV, antibodies to HBc, antibodies to Treponema pallidum and HBsAg.

**CE 2797:** This product is manufactured and distributed in conformity with the Regulation (EU) 2017/746 of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices.

**Batch Release:** The lot has been tested according to valid directions. It meets all required specification criteria and was released for sale.

**Approval:**

18 JUN 2026

Quality Assurance

Date