



Representative Reactivities Table

Optitrol NAT HBV Screening

Positive Control

Version: 1

Date: Jun 12, 2025

Control	Manufacturer	Assay Name	Unit	Mean	SD	Samples	Labs	Representative Reactivity
Optitrol NAT HBV Screening	Roche	cobas® MPX Multiplex HIV, HCV & HBV nucleic acid test for use on cobas® 6800/8800 Systems	IU/mL	48.30	15.0	14	1	positive
	Grifols	Procleix® Ultrio Elite Assay Procleix Panther System	S/CO	23.8	1.2	30	3	positive