

# Revised Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*

Product Description: Seraseq® FFPE WT (DNA/RNA) Reference Material  
 Material Number: 0710-0137 Batch Number: 10587338  
 Material Description: Seraseq® FFPE WT (DNA/RNA) Reference Material contains the reference human cell line GM24385 that have been formalin fixed and embedded in paraffin. The cell line is WT in that it is not derived from cancerous tissue, but mutations from the human reference sequence (GrCh38) are still present.

Date of Manufacture: 13 NOV 2021 Expiration Date: 13 NOV 2024

Fill Volume: One 10 µm curl

Test Method for Concentration: Total nucleic acid was extracted using the Agencourt Formapure Kit and the RNA was quantified using the Qubit RNA HS assay. DNA was extracted using the Qiagen QIAamp DNA FFPE Tissue Kit and quantified using the Qubit dsDNA HS Assay.

Concentration: Average RNA yield per curl: 717 ng (Range 490 ng - 858 ng)  
 Average DNA yield per curl: 335 ng (Range 300 ng - 365 ng)

Test Method for DNA Mutations: ArcherDx VariantPlex Solid Tumor Kit was run on the ILMN MiSeq instrument using 250 ng of DNA. Sequencing depth per sample was 2.3 M reads. Variant analysis was performed using Archer Analysis\_v6.2.7, and "Somatic" Filter Set for variant filtering.

Test Method for RNA Fusions: ArcherDx FusionPlex Solid Tumor Kit run on the ILMN MiSeq instrument using 250 ng of input RNA. Sequencing depth per sample was 2.8 M reads. Data analysis was performed using the Archer Analysis\_v6.2.7 software.

## DNA Sequencing Results:

Gene ID	AA Change	Representative Allele Frequency Result
CDKN2A	p.Asp74Ala	5.11%
ERBB4	Splice variant (c.884-7dup)	15.36%
FOXL2	p.Leu248Pro	18.34%
FOXL2	p.Gln99Lys	10.29%
FOXL2	p.Asn94Lys	3.66%
IDH1	p.Ser9Phe	12.21%
MAP2K1	Splice variant (c.292-3C>T)	51.80%

# Revised Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*

Product Description: Seraseq® FFPE WT (DNA/RNA) Reference Material

RNA Fusion Sequencing Results:

Fusions Detected	Representative Start Sites Result	Representative Reads Result	Representative Percent Total of Reads that Support Fusion Call
COQ8B:NUMBL	16	20	37

Approval:



17 NOV 2023

Prepared By

Date