Revised Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® CNV Breast Mix, +6 copies

Material Number: 0710-0412 Batch Number: 10632778

Material Description: Mixture of human genomic DNA from the reference cell line, GM24385, and synthetic DNA constructs

Concentration (Qubit dsDNA BR Assay): Nominal value: 10 ng/µL
Average measured value: 10.8 ng/µL

Fill Volume: 20 µL

Date of Manufacture: 25 JUL 2022 Expiration Date: 25 JUL 2025

Test Method: BioRad PrimePCR Copy Number Assays run on the BioRad QX-200 Droplet Digital PCR System

Gene	Number of Additional Copies
ERBB2	5.7
FGFR3	5.8
MYC	7.5

Approval:

Prepared By

22 JUL 2024
Date