

# Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*


Product Description:	SeraSeq® Compromised FFPE WT (DNA/RNA) Reference Material		
Material Number:	0710-1710	Batch Number:	10595841
Material Description:	Formalin-fixed paraffin-embedded reference cells (GM24385) manufactured to more closely mimic the quality of patient tissue		
Date of Manufacture:	12 DEC 2021	Expiration Date:	12 DEC 2023
Fill Volume:	One 10 µm curl		
Test Method for DNA Yield per Curl:	DNA was extracted using the Qiagen QIAamp DNA FFPE Tissue Kit and quantified using the Qubit dsDNA HS Assay.		
Average DNA yield per curl:	178 ng (Range 172 ng - 184 ng)		
Test Method for RNA Yield per Curl:	RNA was extracted using the Agencourt FormapureKit and quantified using the Qubit RNA HS Assay.		
Average RNA yield per curl:	1391 ng (Range 1158 ng - 1590ng)		
Test Method for DNA Quality Assessment:	Agilent TapeStation Genomic DNA ScreenTape Analysis		
Average DIN value:	3.95		
Test Method for RNA Quality Assessment:	Agilent Bioanalyzer RNA 6000 Nano Kit Analysis		
Average RIN value:	1.00		
Test Method for Variant Detection:	Archer VariantPlex Solid Tumor and FusionPlex Solid Tumor Kits		
Variant Detection Results:	DNA variants detected:		
	APC p.Val1822Asp COSM3760871 – 51.2% allele frequency		
	KDR p.Gln472His COSM149673 - 51.2% allele frequency		
Variant Detection Results:	KIT p.Met541Leu COSM28026 – 47.7% allele frequency		
	MAP2K1 c.292-3C>T (splice variant) - 51.7 % allele frequency		
	No fusions were detected.		
	Complete results and testing conditions used available upon request.		
Storage:	2 - 8°C		

# Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*

Product Description: Seraseq® Compromised FFPE WT (DNA/RNA) Reference Material

Approval:

A handwritten signature in black ink, appearing to be "EHL".

11 JAN 2022

Prepared By

Date