

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description:	SeraSeq [®] Lymphoma DNA Mutation Mix		
Material Number:	0710-2203	Batch Number:	10800328
Material Description:	A genomic DNA mutation mix of 26 genes typically profiled by targeted NGS assays blended to a background of gDNA extracted from GM24385 (WT) human reference cell line.		
Concentration:	Nominal value: 15 ng/ μ L Average measured value: 18.6 ng/ μ L		
Concentration Test Method:	Qubit ds DNA BR Assay		
dPCR Test Method:	Variant and Wild-type specific PCR assays run on the BioRad QX-200 digital PCR system. Values shown in the table represent the variant allele frequency and are the result of triplicate testing.		
NGS Assay Test Method:	A 250-ng aliquot of the mix was analyzed in triplicate using a custom ArcherDX VariantPlex panel. Default Archer Analysis v6.2.7 settings were applied for calling SNVs and indels. For translocation detection, allele frequency was calculated by dividing the number of observed translocation events by the total observations, which included both the translocation reads and the breakpoint reads from the two corresponding normal genes.		
Fill Volume:	25 μ L		
Date of Manufacture:	20-JAN-2026	Expiration Date:	20-JAN-2030
Storage:	-20°C		

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® Lymphoma DNA Mutation Mix

Variant	Variant Type	% Allele Frequency (dPCR)	% Allele Frequency (NGS)
NPM1-ALK	Translocation	5.1	2.7
HSP90AA1-BCL6	Translocation	6.0	3.0
CCND1-CDC42BPB	Translocation	5.0	2.5
BIRC3-MALT1	Translocation	5.5	3.5
MYC-IGH	Translocation	6.4	3.9
TBL1XR1-TP63	Translocation	5.3	5.3

Gene	Nucleic Acid change	Amino Acid Change	% Allele Frequency (dPCR)	% Allele Frequency (NGS)
BCL2	c.302G>C	p.G101A	6.0	6.6
BRAF	c.1799T>A	p.V600E	5.9	3.7
CXCR4	c.1013C>G	p.S338X	5.3	6.6
CXCR4	c.1013C>A	p.S338X	5.5	5.8
DNMT3A	c.2645G>A	p.R882H	6.2	6.5
EZH2	c.1922A>T	Y641F	5.7	6.9
IDH2	c.515G>A	p.R172K	6.0	7.8
MYD88	c.794T>C	p.L265P	6.0	5.0
NOTCH1	c.7541_7542del	p.P2514Rfs*4	5.9	6.5
NOTCH2	c.7198C>T	p.R2400*	6.1	8.1
RHOA	c.50G>T	p.G17V	5.4	3.9
SF3B1	c.2098A>G	p.K700E	6.0	9.5
STAT3	c.1919A>T	p.Y640F	5.8	7.6
STAT3	c.1982A>T	p.D661V	5.5	5.0
STAT3	c.1940A>T	p.N647I	5.2	4.9
STAT5B	c.1994A>T	p.Y665F	5.3	4.7
STAT5B	c.1924A>C	p.N642H	5.3	5.6
TP53	c.743G>A	p.R248Q	7.4	7.2
TP53	c.820del	p.V274Ffs*71	7.4	6.1
TP53	c.818G>A	p.R273H	6.5	8.0

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® Lymphoma DNA Mutation Mix

Approval:

Prepared By Saima Salim Date 20 Feb 2026