

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® Solid Tumor CNV Mix, +6 copies

Material Number: 0710-2867 Batch Number: 10809087

Material Description: Mixture of human genomic DNA from reference cell line, GM24385, and biosynthetic DNA constructs

Concentration: Nominal value: 10 ng/μL
(Qubit dsDNA BR Assay): Average measured value: 13.9 ng/μL

Fill Volume: 20 μL

Date of Manufacture: 16 MAR 2026 Expiration Date: 16 MAR 2030

Test Method: BioRad PrimePCR Copy Number Assays run on the BioRad QX-200 Droplet Digital PCR System
Archer VariantPlex Solid Tumor Assay run on NextSeq 2000 Sequencing System

Gene	Additional Copies (dPCR)	Additional Copies (NGS)
AKT2	5.9	N/A
BRAF	5.8	N/A
EGFR	5.7	6.3
ERBB2	6.2	6.3
FGFR3	5.8	5.4
KIT	5.6	4.3
KRAS	6.0	3.5
MET	6.4	4.8
MYC	5.7	6.8
MYCN	6.5	4.8
NTRK1	6.6	N/A
PIK3CA	6.1	3.8

Approval:



06 APR 2026

Prepared By

Date