

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® Trisomy 18 Male - Matched Reference Material
Material Number: 0720-0171 Batch Number: 10809780
Material Description: A ccfDNA-like mixture of human genomic DNA (matched maternal and fetal) in a commutable matrix (simulated plasma).
Concentration (Qubit dsDNA BR Assay): 31.0 ng/mL
Volume: 1.0 mL
Date of Manufacture: 10 APR 2026 Expiration Date: 12 SEP 2027
Fetal Fraction by NGS (Chroma Panel): 8.0%

Trisomy Call for Chromosome 18 by NGS (Chroma Panel): Pass
Call for fetal gender by NGS (Chroma Panel): XY
Storage: 2-8°C

Approval:

A handwritten signature in blue ink, appearing to be "M. J. [unclear]".

Prepared By

10 APR 2026

Date